

Call for tenders' details

Title: Supply of an 'in situ' nanomechanical testing system for the JRC Karlsruhe.

Start date: 01/08/2017

Time limit for receipt of tenders: 08/09/2017

Contracting authority: European Commission, Joint Research Centre (JRC)

Status: Closed

Call for tenders question list

#	Submission date	Publication date	Question subject	Question	Answer
1	08/08/2017 09:19	10/08/2017 11:44	Technical Specification 3.5	Does the Site acceptance test have to be executed on both existing (FIB/SEM FEI Versa 3D Dual Beam platform and SEM FEI Quanta 650 model) SEMs or will the test be performed only in one SEM?	10/08/2017 The site acceptance test shall be performed only on the existing FIB/SEM FEI Versa 3D Dual Beam platform.