

## Call for tenders' details

Title: Windowless Automatic Sample Changers for the Measurement of Alpha and Beta Contamination

Start date: 22/04/2021

Time limit for receipt of tenders: 25/05/2021

Contracting authority: European Commission, Joint Research Centre - Geel (JRC-GEE)

Status: Closed

## Call for tenders question list

#	Submission date	Publication date	Question subject	Question	Answer
1	12/05/2021 10:35	12/05/2021 11:31	Tritium detection requirement	Our system doesn't allow capability to detect soft beta radiation from Tritium as requested in the detector specification of document EN-1.2. Technical Specifications. Will our offer therefore be disqualified or isn't this requirement mandatory?	<b>12/05/2021</b> The technical specifications are mandatory technical requirements. Detecting soft beta radiation from Tritium is the most important requirement for this system. If the offer does not fulfil all mandatory requirements, it will be disregarded for the evaluation.

Generated on the 24/04/2024 05:05:10 - Generation time 1 ms